www.nature.com/ejcn

ORIGINAL ARTICLE Clinical anthropometrics and body composition from 3D whole-body surface scans

BK Ng^{1,2}, BJ Hinton^{1,2}, B Fan¹, AM Kanaya³ and JA Shepherd^{1,2}

BACKGROUND/OBJECTIVES: Obesity is a significant worldwide epidemic that necessitates accessible tools for robust body composition analysis. We investigated whether widely available 3D body surface scanners can provide clinically relevant direct anthropometrics (circumferences, areas and volumes) and body composition estimates (regional fat/lean masses).

SUBJECTS/METHODS: Thirty-nine healthy adults stratified by age, sex and body mass index (BMI) underwent whole-body 3D scans, dual energy X-ray absorptiometry (DXA), air displacement plethysmography and tape measurements. Linear regressions were performed to assess agreement between 3D measurements and criterion methods. Linear models were derived to predict DXA body composition from 3D scan measurements. Thirty-seven external fitness center users underwent 3D scans and bioelectrical impedance analysis for model validation.

RESULTS: 3D body scan measurements correlated strongly to criterion methods: waist circumference $R^2 = 0.95$, hip circumference R^2 = 0.92, surface area R^2 = 0.97 and volume R^2 = 0.99. However, systematic differences were observed for each measure due to discrepancies in landmark positioning. Predictive body composition equations showed strong agreement for whole body (fat mass R^2 = 0.95, root mean square error (RMSE) = 2.4 kg; fat-free mass R^2 = 0.96, RMSE = 2.2 kg) and arms, legs and trunk (R^2 = 0.79-0.94, RMSE = 0.5–1.7 kg). Visceral fat prediction showed moderate agreement (R^2 = 0.75, RMSE = 0.11 kg).

CONCLUSIONS: 3D surface scanners offer precise and stable automated measurements of body shape and composition. Software updates may be needed to resolve measurement biases resulting from landmark positioning discrepancies. Further studies are justified to elucidate relationships between body shape, composition and metabolic health across sex, age, BMI and ethnicity groups, as well as in those with metabolic disorders.

European Journal of Clinical Nutrition (2016) 70, 1265–1270; doi[:10.1038/ejcn.2016.109;](http://dx.doi.org/10.1038/ejcn.2016.109) published online 22 June 2016

INTRODUCTION

Regional body shape and composition provide stronger indicators of obesity-related metabolic risk than the body mass index (BMI). Goodpaster et $al¹$ $al¹$ $al¹$ showed that normal weight men with high visceral adipose levels were twice as likely to have metabolic syndrome. Wilson et al^2 al^2 showed that the individuals in the highest quintile of trunk-to-leg volume ratio are at 6.8 times greater risk for diabetes. Common methods for body composition assessment include bioelectrical impedance analysis (BIA), air displacement plethysmography (ADP) and dual energy X-ray absorptiometry (DXA). Of these, DXA is the only method that provides regional information, but DXA is not suitable for large populations because of its relatively high cost and use of ionizing radiation. There is an unmet need for accessible tools for accurate regional body composition assessment.

We investigated the use of 3D whole-body surface scanning for clinical anthropometry, as well as total and regional body composition measurement. These systems generate surface renderings and automated circumference and length measurements across the entire body. Several studies have assessed their ability to accurately and precisely quantify clinically relevant measures. Wells et al^3 al^3 reported 0.5 cm precision on body circumferences (chest, waist, hips and so on), using a six-camera structured light scanner. Wang *et al.^{[4](#page-5-0)}* reported precise (coefficient of variation (%CV) = 0.38) whole-body volume measurements using a four-camera lasertriangulation scanner. Lin et $al^{5,6}$ $al^{5,6}$ $al^{5,6}$ reported correlations between 3D anthropometric measures and metabolic risk factors. Lee et al ^{[7](#page-5-0)} demonstrated accurate prediction of whole-body and regional fat mass and percent fat from regional volume and length measures from an eight-camera structured illumination scanner. Development of low-cost light-coding technology has enabled more affordable 3D scanners such as the TC^2 KX-16 (Cary, NC, USA) 8 and the Fit3D Proscanner (Redwood City, CA, USA).

In summary, 3D body surface scanning is a compelling tool for metabolic status assessment that offers inexpensive, radiation-free and automated collection of hundreds of measurements that would otherwise require significant time and personnel resources to collect. The objective of this study was to validate direct anthropometric and derived body composition measurements from 3D whole-body surface scans against criterion methods.

METHODS

We conducted a cross-sectional stratified study of healthy adults. 3D scan measurements (circumferences, surface areas and volumes) were systematically compared with manual anthropometry, DXA and ADP. Predictive equations were derived to estimate DXA body composition using 3D scan measurements.

Received 23 December 2015; revised 29 April 2016; accepted 23 May 2016; published online 22 June 2016

¹Department of Radiology and Biomedical Imaging, University of California, San Francisco, San Francisco, CA, USA; ²The UC Berkeley - UCSF Graduate Program in Bioengineering, San Francisco, CA, USA and ³Department of Medicine, University of California, San Francisco, San Francisco, CA, USA. Correspondence: Professor JA Shepherd, Department of Radiology and Biomedical Imaging, University of California, San Francisco, MRSC A-C108, 1 Irving Street (Box 0628), San Francisco, CA 94143, USA. E-mail: john.shepherd@ucsf.edu

1266

Participants

There were two participant groups: a calibration group and a field validation group. Calibration participants were recruited using flyers posted around UCSF between January 2014 and May 2015. Calibration group recruitment was stratified by age (20–40, 40–60, 60+ years), sex and BMI (normal $<$ 25 kg/m², , 25 kg/m² \geq overweight < 30 kg/m², obese \geq 30 kg/m²). Eligible participants were identified as ambulatory individuals within the study strata. Exclusion criteria included current pregnancy, missing limbs, non-removable metal in the body (for example, joint replacements), a history of body-altering surgery (for example, liposuction) and significant hair that cannot be contained within a swim cap. Each calibration participant underwent a whole-body DXA scan and two 3D optical scans, with repositioning. Weight, height, and waist and hip circumferences (each in duplicate) were recorded. All measures were acquired at the UCSF Clinical and Translational Science Institute (CTSI) Body Composition, Exercise Physiology and Energy Metabolism Laboratory (San Francisco, CA, USA). Calibration participants gave informed consent, and the study protocol was approved by the UCSF Committee on Human Research.

Participants in the field validation group were scanned at one of the eight fitness centers in the United States and Australia for body shape selfassessment purposes between July and November 2014. Each participant had a BIA percent fat measurement and a 3D body scan on the same day. Validation participants were selected to have height, weight, BMI and age measurements within the minimum-maximum ranges of the calibration data set. It should be noted that BIA measurements were user-reported, and the BIA hardware and acquisition protocols were uncontrolled. All field validation participants signed a waiver of consent authorizing the use of their anonymized scans by the investigators.

3D surface scans

3D surface scans were acquired on a Fit3D Proscanner (Fit3D Inc., Redwood City, CA, USA) according to a standard protocol. The device consists of a rotating platform and a 3D optical light-coding camera mounted in a tower 2 m from the center of the platform. Users grasp adjustable handles mounted on the platform such that their arms are straight and relaxed, abducted from the body. When buttons on the scanner handles are depressed by the user, a 360-degree 3D image is acquired while the platform rotates once around in approximately 40 s. Although only 11 circumferences (chest, waist and hips, as well as left/right biceps, forearm, thigh and calf) are reported to end users, 476 anthropometric measurements from the neck down to the ankles and wrists are automatically derived and stored in a proprietary database. In general, the head, hands and feet are excluded from all circumference, surface area and volume calculations. The Fit3D system was chosen for this study over other models because approximately 100 of these systems are available to the public at fitness centers across the US.

Each calibration participant was scanned twice, with repositioning. Formfitting boxer briefs and a swim cap were provided for each participant. Male participants were scanned topless, whereas female participants wore a sports bra. Each validation participant was scanned once, in personal form-fitting clothing, with long hair tied above the neck. 3D scan data and measurements were transferred securely from Fit3D to UCSF.

For quality control, we performed 41 scans of a female mannequin (part #DSPEFAMW, Display Warehouse, San Diego, CA, USA) with repositioning over a 2-month period (data not included). Chest, hip, thigh and waist circumference, as well as total body volume, showed high long-term stability with CV between 0.25 and 1.2%.

Abbreviations: BMI, body mass index; FFMI, fat free mass index; FMI, fat mass index. Reported circumferences, areas and volumes were derived from the 3D optical scans. Unpaired two-tailed t-tests were performed to detect significant mean differences in the direct 3D measurements between the two groups (P-values shown in the right-most column). Note that body composition was measured using DXA on the calibration group and BIA on the validation group. ^a ^adenotes $P \le 0.05$.

Clinical 3D anthropometry and body composition BK Ng et al

Figure 1. 3D scanner body circumferences and surface area vs criterion methods. Clockwise from top left: hip circumference, waist circumference and body surface area comparisons, then sample matched DXA and 3D optical images with annotated landmarks. 3D circumferences and surface area show high correlation with manual tape measures. Biases may be the result of non-identical landmark positioning between the methods. Of note, body surface area shows significant bias that may be explained by the fact that the 3D scanner software does not report surface area of the head and neck. This omission is not a technical shortcoming but a design decision by the manufacturer to avoid inaccuracies introduced by voluminous hair.

Dual energy X-ray absorptiometry

Only calibration participants received a DXA scan as the criterion method for body composition. Whole-body scans were acquired on a Hologic Discovery/W or Horizon/A system (Hologic Inc., Malborough, MA, USA). All scans were centrally analyzed at UCSF by a single ISCD-certified technologist using Hologic Apex software (version 13.5.2.1) and NHANES calibration.^{[9](#page-5-0)} Participants were scanned in examination gowns, without shoes. Participants were centered on the scanner table with arms out to the side, hands flat on table and feet in planarflex position, in accordance with the manufacturer's standard protocols.¹⁰ The standard DXA output includes percent fat, fat mass, lean mass and total mass for the whole body, as well as the arms, legs, torso and head. In addition, total and regional body volumes were derived for each region on the DXA report (arms, legs, trunk, whole body) using the equations of Wilson et al .^{[10](#page-5-0)}

Air displacement plethysmography

Only calibration participants received an ADP measurement as the criterion method for total body volume. Measurements were performed using a Bod Pod (COSMED USA, Inc., Chicago, IL, USA). Before each measurement, the instrument was calibrated by placing a hollow cylinder with known volume into the Bod Pod. Participants wore the same outfit for the ADP measure as for the 3D optical scan (swim cap, boxer briefs and a sports bra for females). Body volume is measured by the body's air displacement with corrections for residual lung volume and surface area artifacts. Details of the standard ADP protocol are described elsewhere.¹¹

Statistical methods

All analyses were performed using SAS version 9.4 (SAS Institute, Cary, NC, USA). Univariate linear regressions were performed to assess agreement of selected clinically relevant anthropometric measurements acquired on the 3D scanner versus criterion methods: hip and waist circumference tape measurements, body surface area estimated using the Du Bois model¹² and whole/regional body volumes calculated from ADP and DXA, respectively. Student's t-tests were performed to detect significant measurement biases between 3D measurements and criterion methods. Measurements were assumed to be normally distributed. A critical P-value of 0.05 was considered significant. %CV and root mean square error (RMSE) were calculated for the matched test-retest measurements from the 3D optical scanner. Coefficients of determination (R^2) reported in this study are adjusted for multiple variables where applicable.

Predictive equations were derived for whole-body and regional DXA body composition variables, including percent fat, fat mass and visceral fat mass. Where available, we derived linear regression equations using the parameters described by Lee et $al^{7,13}$ $al^{7,13}$ $al^{7,13}$ The ratio of waist girth to waist width was used as a surrogate for 'central obesity depth' defined by Lee et al. Waist girth and width were likewise used as surrogates for 'central obesity width' as defined by Lee et al. We further derived predictive equations for

1268

Measurement type	Variable	%CV	RMSE
Circumference (cm)	Waist	1.50 1.41	
	Hips		0.75 0.79
	Biceps		2.24 0.77
	Forearm		1.93 0.54
	Thigh		0.95 0.59
	Calf		0.92 0.36
Surface area $(m2)$	Whole body		1.38 0.0168
	Torso	0.81	0.0057
	Average (L/R) arm		3.45 0.0044
	Average (L/R) leg		2.75 0.0074
Measured volume (I)	Whole body	0.74 0.57	
	Average (L/R) arm	4.49 0.16	
	Average (L/R) leg	2.61	0.23
	Trunk	0.99 0.51	
Derived fat/fat-free mass (kg)	Whole-body fat		1.96 0.50
	Whole-body fat-free	0.94 0.50	
	Whole-body percent fat (%)		2.16 0.68
	Visceral fat mass	6.69 0.03	
	Trunk fat		2.38 0.30
	Trunk fat-free	0.50 0.13	
	Arms fat	11.63	0.34
	Arms fat-free	6.67 0.42	
	Legs fat		1.25 0.11
	Legs fat-free		1.99 0.36

articipant in the calibration data set was scanned twice, with repositioning.

fat-free mass in each compartment (whole body, legs and trunk), using the same parameters as the fat mass equations. Equations were derived using linear regression (proc GLM). Additional equations were derived for fat and lean mass in the arms, by using stepwise linear regression (proc GLMSELECT, selection = STEPWISE) and over 18 length, area and volume measurements of the arms and trunk. Selection was performed subject to minimization of the Schwarz-Bayesian Information Criterion, with fivefold cross validation on the calibration data.

As DXA scans were acquired on two different systems (Horizon/A and Discovery/W), we performed covariate analysis on each equation to determine whether cross-calibration was necessary. Adjusting for age, height and weight, no significant differences were found between the two scanners for all predicted variables, except visceral fat. Consequently, all measurements, except visceral fat, were pooled directly between the two scanners. Visceral fat measurements acquired on the Discovery/W were calibrated to the Horizon/A, using a linear calibration equation derived from a separate study of 13 participants who underwent sequential whole-body scans on the two scanners (unpublished work).

Whole-body prediction equations were applied to the field validation data. Estimated whole-body percent fat, fat mass and fat-free mass were compared with the reported values from bioimpedance analysis using simple linear regression.

RESULTS

Thirty-nine individuals (20 male) completed the calibration study. Thirty-seven individuals (18 male) were included in the validation group. Summary statistics of the calibration and validation populations are provided in [Table 1.](#page-1-0) Example 3D optical and DXA images are shown in [Figure 1](#page-2-0).

Regression plots are shown in [Figure 1](#page-2-0) for 3D optical measurements against tape measure circumferences and Du Bois-estimated body surface area. [Figure 2](#page-4-0) shows 3D measurements against DXA and ADP whole-body volumes, and DXA regional volumes. Strong association was observed for waist and hip circumferences $(R^2 = 0.95$ and 0.92, respectively). T-tests showed significant mean differences of 1.75 cm (95% confidence interval: (0.58, 2.91)) for waist circumference and 3.17 cm (95% confidence interval: (1.93, 4.41)) for hip circumference between the 3D scanner and tape measurements.

1269

Figure 2. 3D scanner body volume vs criterion methods. Clockwise from top left: total volume vs ADP and DXA, arms vs DXA, legs vs DXA and trunk vs DXA. High precision and accuracy is observed for whole-body volume. Because of differences in partition locations, regional measurements are not directly comparable between the 3D scanner and DXA. For instance, the legs are partitioned horizontally at the crotch in the 3D optical system but diagonally from the crotch up to the top of the hip in the DXA system. Consequently, reported leg volumes are lower on the 3D system than the DXA system. Refinement of the regional partitions is necessary before cross-modality comparisons can be performed.

Surface area and volume measurements from the 3D scanner showed high test-retest precision [\(Table 2](#page-3-0)). Strong association to the Du Bois model was observed for whole-body surface area (R^2 = 0.97), although this 3D system significantly underreports surface area (mean difference -0.38 m², 95% confidence interval: $(-0.40, -0.36)$). Similarly, strong associations to ADP- and DXA-measured wholebody volumes were observed (R^2 = 0.99 and 0.97, respectively), with a significantly smaller 3D scan measured volume relative to ADP (mean difference − 4.15 l, 95% confidence interval: (−5.13, − 3.17)). Regional 3D scan volume estimates were highly correlated to similar measures derived from the DXA scans ($R^2 = 0.73 - 0.97$). In general, the 3D scanner includes less volume in the arm and leg compartments than DXA and correspondingly more volume in the trunk compartment (all $P < 0.001$).

The derived body composition equations for percent fat, fat mass, lean mass and visceral fat mass are shown in [Table 3](#page-3-0). Validation results for the whole-body measurements are included. Fat and fat-free mass models exhibited strong fit to the calibration data (R^2 = 0.95 and 0.96) and a reasonable fit to the validation data $(R²=0.76$ and 0.85). The visceral fat prediction equation showed moderately strong association (R^2 = 0.75). Predictive models for regional fat and lean mass showed generally strong associations

estimates was validated in a separate data set using BIA. Although some biases were found in the anthropometric measures, this

DISCUSSION

study supports the use of 3D surface scanning as an accurate, precise and automated substitute to other methods such as measuring tape, ADP and DXA.

on the calibration set ($R^2 = 0.79 - 0.94$); however, no validation data

In this study of automated anthropometric measures from 3D whole-body surface scans, we found strong associations of waist and hip circumference to tape-measured values, body surface area to the Du Bois model and body volumes to DXA volume estimates. 3D measures were used to derive whole-body and regional body composition estimates for both sexes across a wide range of ages and BMI values. The accuracy of these body composition

were available for these regional models.

Notably, 3D scanning is a more direct measure of surface area than the criterion method (height/weight equation) available. In practice, there is no established gold standard for body surface area measurement. We found strong correlation between surface area measurements from the present 3D scanner and the clinically prevalent Du Bois model ($R^2 = 0.97$). Tikuisis et al.^{[14](#page-5-0)} reported high

1270

precision and accuracy of a 3D laser scanner to six different height and weight equations for body surface area, but again no gold standard comparison method was available. This measure is clinically relevant for modeling evaporative water loss, in particular for burn injuries 15 and calculating chemotherapeutic medication dosages.¹⁶

Whole-body volume measurements from 3D scans exhibited high precision (%CV = 0.74) comparable to ADP (%CV = 0.10).¹⁷ Differences in landmark and partition positioning in the 3D surface scan analysis algorithms led to significant biases in regional volume measurements compared with DXA. For example, the present trunk/ leg partition defined by the Proscanner is a horizontal plane at the crotch, whereas on a standard Hologic DXA, the trunk/leg partitions run diagonally from the crotch to the hips, such that the femur is completely in the leg compartment. Calibration to DXA compartments would enable direct assessment of trunk-to-leg volume ratio, 2 a strong independent indicator of metabolic health and disease risk.

Body composition models from 3D features calibrated using DXA data validated well, similar to those reported by Lee et $al.7^{7,18}$ especially in light of the different body composition measurement methods used (DXA calibration data and BIA validation data in our study and DXA and MRI data in Lee's studies). In particular, our whole-body fat mass prediction model showed strong fit to the calibration data (R^2 = 0.95), matching the equivalent Lee model $(R^2 = 0.95)$.⁷ Our visceral fat prediction model showed moderately strong association (R^2 = 0.75), similar to Lee (R^2 = 0.72).¹³

This study had a few limitations. Small sample size ($n = 39$) limits the statistical power of our models. However, all age/BMI/sex strata defined for this study were represented. Another issue was that the validation data contained only BIA body composition data rather than DXA data. These BIA data were reported by users rather than trained technicians with a defined protocol. Despite these limitations, the strong validation results suggest good predictive model stability across a range of body shapes in a real consumer environment. Finally, our findings were derived from a healthy population without any known conditions that may alter the relationship between 3D body shape and body composition such as sarcopenia, anorexia or malnutrition.

We conclude that 3D surface scanning presents a compelling modality for clinical anthropometry. This method provides an accessible platform for rapid body measurement, as well as total and regional body composition analysis. Because of low cost, high precision and a lack of ionizing radiation, 3D surface scanning is uniquely suitable for routine clinical use to monitor longitudinal metabolic health. This study shows feasibility of broad clinical use of 3D surface scanning to estimate body composition in a wide range of body shapes. Larger follow-up studies are justified to better understand the relationships of 3D body shape and composition across various sex, BMI, age and ethnicity groups, as well as in special populations with metabolic conditions and potentially abnormal body composition.

CONFLICT OF INTEREST

The authors declare no conflict of interest.

ACKNOWLEDGEMENTS

We acknowledge Dr Kathleen Mulligan for fostering collaboration with the UCSF Clinical and Translational Science Institute, Viva Tai and Caitlin Sheets for guiding participants through the clinical DXA protocols, as well as Leila Kazemi, Louise Marquino and Eboni Stephens for their contributions as study coordinators. We also thank Greg Moore and Tyler Carter of Fit3D Inc. for their support with the Fit3D Proscanner and processing of 3D scan data.

AUTHOR CONTRIBUTIONS

JAS and BKN designed and conducted the research; BKN, BJH and JAS analyzed data; BKN and JAS drafted the manuscript and had primary responsibility for final content. All authors reviewed and approved the final manuscript.

REFERENCES

- 1 Goodpaster BH, Krishnaswami S, Harris TB, Katsiaras A, Kritchevsky SB, Simonsick EM et al. Obesity, regional body fat distribution, and the metabolic syndrome in older men and women. Arch Intern Med 2005; 165: 777–783.
- 2 Wilson JP, Kanaya AM, Fan B, Shepherd JA. Ratio of trunk to leg volume as a new body shape metric for diabetes and mortality. PLoS ONE 2013; 8: e68716.
- 3 Wells JC, Treleaven P, Cole TJ. BMI compared with 3-dimensional body shape: the UK National Sizing Survey. Am J Clin Nutr 2007; 85: 419–425.
- 4 Wang J, Gallagher D, Thornton JC, Yu W, Horlick M, Pi-Sunyer FX. Validation of a 3-dimensional photonic scanner for the measurement of body volumes, dimensions, and percentage body fat. Am J Clin Nutr 2006; 83: 809-816.
- 5 Lin JD, Chiou WK, Weng HF, Fang JT, Liu TH. Application of three-dimensional body scanner: observation of prevalence of metabolic syndrome. Clin Nutr 2004; 23: 1313–1323.
- 6 Lin JD, Chiou WK, Weng HF, Tsai YH, Liu TH. Comparison of three-dimensional anthropometric body surface scanning to waist-hip ratio and body mass index in correlation with metabolic risk factors. J Clin Epidemiol 2002; 55: 757–766.
- 7 Lee JJ, Freeland-Graves JH, Pepper MR, Stanforth PR, Xu B. Prediction of android and gynoid body adiposity via a three-dimensional stereovision body imaging system and dual-energy X-ray absorptiometry. J Am Coll Nutr 2015; 2015: 1-11.
- 8 Daanen HAM, Ter Haar FB. 3D whole body scanners revisited. Displays 2013; 34: 270–275.
- 9 Schoeller DA, Tylavsky FA, Baer DJ, Chumlea WC, Earthman CP, Fuerst T et al. QDR 4500 A dual-energy X-ray absorptiometer underestimates fat mass in comparison with criterion methods in adults. Am J Clin Nutr 2005; 81: 1018-1025.
- 10 Wilson JP, Mulligan K, Fan B, Sherman JL, Murphy EJ, Tai VW et al. Dual-energy X-ray absorptiometry–based body volume measurement for 4-compartment body composition. Am J Clin Nutr 2012; 95: 25–31.
- 11 Fields DA, Goran MI, McCrory MA. Body-composition assessment via air-displacement plethysmography in adults and children: a review. Am J Clin Nutr 2002; 75: 453–467.
- 12 Du Bois D, Du Bois EF. Clinical calorimetry: Tenth paper a formula to estimate the approximate surface area if height and weight be known. Arch Intern Med 1916; XVII: 863–871.
- 13 Lee JJ, Freeland‐Graves JH, Pepper MR, Yao M, Xu B. Predictive equations for central obesity via anthropometrics, stereovision imaging and MRI in adults. Obesity 2014; 22: 852–862.
- 14 Tikuisis P, Meunier P, Jubenville C. Human body surface area: measurement and prediction using three dimensional body scans. Eur J Appl Physiol 2001; 85: 264–271.
- 15 Jelenko C 3rd. Studies in burns. I. Water loss from the body surface. Ann Surg 1967; 165: 83.
- 16 Pinkel D. The use of body surface area as a criterion of drug dosage in cancer chemotherapy. Cancer Res 1958; 18: 853–856.
- 17 Collins A, McCarthy H. Evaluation of factors determining the precision of body composition measurements by air displacement plethysmography. Eur J Clin Nutr 2003; 57: 770–776.
- 18 Lee JJ, Freeland-Graves JH, Pepper MR, Yu W, Xu B. Efficacy of thigh volume ratios assessed via stereovision body imaging as a predictor of visceral adipose tissue measured by magnetic resonance imaging. Am J Hum Biol 2015; 27: 445–457.